

Caliper Equipment Description

Equipment ID: _____
Manufacturer: _____
Model #: _____
Serial #: _____
Storage Location: _____

Description: Caliper – readable to 0.01 in (0.2 mm) or better

Primary Purpose: Measuring distances

Calibration Item: Verify measured distances

Calibration Procedure: In-House Procedure _____
Calibration Worksheet

Calibration Equipment: NIST traceable calibration block(s)

Calibration Interval: 12 months

Specifications: AASHTO T 22
ASTM C39

Specified Tolerance: As stated in the applicable test method