

## Mallet Equipment Description

Equipment ID: \_\_\_\_\_  
Manufacturer: \_\_\_\_\_  
Model #: \_\_\_\_\_  
Serial #: \_\_\_\_\_  
Storage Location: \_\_\_\_\_

Description: Rubber or rawhide head mallet

Primary Purpose: Releasing trapped air from fresh concrete after consolidation

Calibration Item: Verify mass of mallet

Calibration Procedure: In-House Procedure \_\_\_\_\_  
Calibration Worksheet

Calibration Equipment: Scale readable to 0.01 lb. (1 g) or better

Calibration Interval: 12 months

Specifications: AASHTO T 23 (ASTM C 31)

Specified Tolerance: As stated in the applicable test method