

Neoprene Pad Equipment Description

Equipment ID: _____
Manufacturer: _____
Model #: _____
Serial #: _____
Storage Location: _____

Description: Neoprene Pad

Primary Purpose: Determination of compressive strength of cylinders

Calibration Item: Verify critical dimensions and document usage

Calibration Procedure: In-House Procedure _____
Calibration Worksheet

Calibration Equipment: Ruler or caliper readable to at least 1/16 inch or 1 mm

Calibration Interval: Prior to first use and daily usage documentation

Specifications: AASHTO T 22
ASTM C1231

Specified Tolerance: As stated in the applicable test method