

## Splitter Equipment Description

Equipment ID:  
Manufacturer:  
Model #:  
Serial #:  
Storage Location:

Description: Coarse Aggregate Mechanical Sample Splitter

Primary Purpose: Splitting aggregate samples to testing size

Calibration Item: Verify Critical Dimensions and Check Physical Condition

Calibration Procedure: In-House Procedure  
Calibration Worksheet

Specified Tolerance: As stated in AASHTO T-248

Calibration Equipment: Calipers or Ruler, readable to 1 mm

Calibration Interval: 12 months

Specifications: AASHTO T-248