

## Strike-off Plate Equipment Description

Equipment ID: \_\_\_\_\_  
Manufacturer: \_\_\_\_\_  
Model #: \_\_\_\_\_  
Serial #: \_\_\_\_\_  
Storage Location: \_\_\_\_\_

Description: Strike-off Plate

Primary Purpose: Leveling the surface of fresh concrete flush with the rim of a measure

Calibration Item: Verify critical dimensions and planeness of edges

Calibration Procedure: In-House Procedure \_\_\_\_\_  
Calibration Worksheet

Calibration Equipment: Ruler readable to 1/16 inch (1 mm)  
Straightedge at least 12 in. (300 mm) in length  
Hex Key – 1/16 in. (1.5 mm)

Calibration Interval: 12 months

Specifications: AASHTO T 121, AASHTO T 152  
ASTM C138, ASTM C231

Specified Tolerance: As stated in the applicable test method