

Temperature Recorder Equipment Description

Equipment ID: _____
Manufacturer: _____
Model #: _____
Serial #: _____
Storage Location: _____

Description: Temperature Recorder

Range: To encompass the applicable range of temperatures used

Accuracy: ± 2 °F (± 1 °C)

Primary Purpose: Recording the temperature of liquid baths or moist rooms constantly or at least every 15 minutes

Calibration Item: Verify temperature reading of temperature recorder

Calibration Procedure: In-House Procedure _____
Calibration Worksheet

Calibration Equipment: Reference Thermometer – NIST traceable, readable to 0.5 °C (1 °F)

Calibration Interval: 6 months

Specifications: AASHTO M 201, ASTM C511

Specified Tolerance: As stated in the applicable specification