

## IN-HOUSE PROCEDURE # \_\_\_\_\_

### PROCEDURE FOR VERIFYING LEVELING PLATE FOR AC GAUGE

Item:

Leveling Plate for AC Gauge

Purpose:

This method provides instructions for checking the critical dimensions of a leveling plate used with the nuclear asphalt content gauge

Inspection Equipment Required:

1. Ruler readable to 1/16 inch (1 mm) or better.

Tolerance:

Equipment shall meet the dimensional tolerances specified in the applicable test method.

Procedure:

1. Measure the length of the leveling plate.
2. Measure the width of the leveling plate.
3. Measure the thickness of the leveling plate.