Coarse Sieve Calibration Worksheet

Equipmer	nt ID:			Date:		
Manufact	urer:		Perfori	med By:		
Mod	 lel #:		Last Cal			
Serial #:			Next Calibration Due:			
			Next Calibrati	on Due		
Storage Loca	ition:					
	and	ify physical condition			4 (4.75 m	ım)
Calibration F		House Procedure for	, ,	e Sieves		
Calibration Equipment: Caliper readable to at least 0.01 mm						
	Cal	iper I.D.				
Sieve Opening	Size:		Round Rect	angular Fra	ame Size:	
	□ 0'···· 0'		Condition	S		
Label	☐ Sieve Si			Serial Number		
Frame	☐ Excellen	-			□ Poor	
Joint	☐ Excellent ☐ Good					
Cloth Uniformity Excellent					□ Pc	
Cloth Tightness	☐ Excellen	t 🗆 Good	I 🗆 F	-air	☐ Po	or
Opening	Horizontal	Vertical	Tolerance	(± X)	Pass /	
1					Pass	□ Fail
2					Pass	☐ Fail
3					Pass	□ Fail
4					Pass	□ Fail
5					Pass	☐ Fail
6					Pass	☐ Fail
7					Pass	☐ Fail
8					Pass	☐ Fail
			Tolerance		Pass /	
Avg. Opening					Pass	☐ Fail
Table 1, ASTM E	11 (condensed) U.S. Alternative	± Y Variation for	± X Max Variation	1		
Size (mm)	Sieve Size	Avg. Opening (mm)	for Opening (mm)	1		
50	2 in.	1.34	2.06	Bass	/ Foil	
37.5	1 ½ in.	1.01	1.67	j Pass	/ Fail —	
25 19	1 in. ³ / ₄ in.	0.682 0.522	1.24 1.01	Init	ial By	
12.5	1/ ₂ in.	0.346	0.75	1		
9.5	3/ ₈ in.	0.265	0.61	1		
4.75	No. 4	0.135	0.37	†		

