

Cone and Tamper Calibration Worksheet

Equipment ID: _____ Date: _____
Manufacturer: _____ Performed by: _____
Model #: _____
Serial #: _____ Next Calibration Due: _____
Storage Location: _____ Last Calibration: _____

Calibration Item: Verify critical dimensions of the cone and tamper

Calibration Procedure: In-House Procedure

Calibration Equipment: Calipers, Scale CL1, and ruler read. to 1 mm.

Top Inside Diameter of Cone: _____	_____ Tolerance = 40 +/- 3 mm
Bottom Inside Diameter of Cone: _____	_____ Tolerance = 90 +/- 3 mm
Height of Cone: _____	_____ Tolerance = 75 +/- 3 mm
Thickness of Cone: _____	_____ Tolerance = at least 0.8 mm

Diameter of tamper face: _____	_____ Tolerance = 25 +/- 3 mm
Mass of tamper: _____	_____ Tolerance = 340 +/- 15 g
	Pass/Fail: _____

Comments: _____

Initial By: _____