Sulfate Soundness Temperature Recorder Calibration Worksheet

Equipment ID:	Date:	
Manufacturer:	Performed By:	
Model #:	Last Calibration:	
Serial #:	Next Calibration Due:	
Storage Location:	Temperature Range:	

Calibration Item: Verify temperature reading of water tank temperature recorder

Calibration Procedure: In-House Procedure for Verifying Temperature Recorder

Calibration Equipment: Reference thermometer – NIST traceable, readable to 0.1 °C (0.2 °F)

Manufacturer

Model

Serial No.

	Reference	Temperature	Tolerance	Pass / Fail		
Date	Thermometer	Recorder	± 0.3 ° C			Performed By
	Reading	Reading	± 0.5 °F			
				Pass	🗆 Fail	
				Pass	🗆 Fail	
				Pass	🗆 Fail	
				Pass	🗆 Fail	
				Pass	🗆 Fail	
				Pass	🗆 Fail	
				Pass	🗆 Fail	
				Pass	🗆 Fail	
				Pass	🗆 Fail	
				Pass	🗆 Fail	
				Pass	🗆 Fail	
				Pass	🗆 Fail	
				Pass	🗆 Fail	
				Pass	🗆 Fail	
				Pass	🗆 Fail	
				Pass	🗆 Fail	
				Pass	🗆 Fail	
				Pass	🗆 Fail	
				□ Pass	🗆 Fail	
				Pass	🗆 Fail	

Initial By _____



This document has been created by CTTP. The information provided herein is for educational purposes and does not necessarily meet requirements of other organizations.